



HCELL1 with optional Pt flag type counter electrode

The H-Type Cell, HCELL1, consists of two glass cell which can be interconnected by a membrane or a permeation foil, which should be prepared by user. Each cell chamber may be filled with a different solution and they are linked together by a clamp. The volume of each cell is 1 liter and all wetted parts are made of chemically resistant materials such as Teflon, Pyrex and SUS 316.

The "standard" cell configuration consists of a flat specimen holder for working electrode, a graphite rod counter electrode or platinum flag type electrode, a reference electrode in a Luggin capillary and a gas bubbling tube.

The graphite rod counter electrode, platinum flag type electrode, flat specimen holder and a reference electrode should be ordered separately.

Specifications

Dimensions

Cell Vial Volume	1 liter maximum * 2ea
Sample Volume	about 750 ml / each vial
Size	Approximately 14 cm diameter by 30 cm high including electrodes and adapters

Cylindrical Sample Holder Material

Tube	Pyrex, Ø1/4"
Compression Gasket	Teflon®
Cylindrical Metal Sample	Steel

Chemical Compatibility

Wetted Materials	Pyrex®, Teflon®
Non-Wetted Materials	Above, plus stainless steel and Viton®

Graphite Counter Electrode (option)

Type	High density graphite rod
Size	6 mm diameter by 30 cm length

Pt Flag Type Counter Electrode (option)

Pt Plate	Active Area : 5 cm ² or 25 cm ² , 0.2mm thickness
Length	150mm
Glass Tube Diameter	6mm
Rod Material	Brass

Reference Electrode (option)

Type	SCE or Ag/AgCl reference electrode
Size	9 mm diameter OD & 4mm diameter OD

Flat Specimen Holder (option)

Specimen Diameter	15.5mm ~ 22mm
Specimen Thickness	0.3~5.8mm

Parts Included


1. Cell Vial	Pyrex®, O-ring type beaker, 1L, Ø103(OD)*157(H), 2ea
2. Cell Clamp	Stainless steel
3. Clamp for connecting the half cells	Stainless steel
4. Luggin Capillary	Pyrex®
5. Bubbler Adapter Gas Inlet	Pyrex®
6. Cell Top	Teflon®
7. Cylindrical Metal Sample & Tube	Steel / Quartz tube, Ø6.35* Ø4.35
8. Other miscellaneous such as stopper / O-ring	MC Nylon® / Viton®

Optional Items

1. Flat Specimen Holder	FSH2
2. Graphite Counter Electrode	GR002
3. Pt Flag Type Counter Electrode	PFL5 or PFL25 depending on active area (5cm ² or 25cm ²)
4. Ag/AgCl Reference Electrode	1004
5. Saturated Calomel Reference Electrode	1001

(Please contact us for other replacement parts.)

Cell Top



Flat Specimen Holder, FSH2



Pt Flag Type Electrode, PFL



**Saturated Calomel Reference Electrode, 1001
Ag/AgCl Reference Electrode, 1004**



Graphite Rod, GR002



WonATech Co., Ltd.

Main Office & R&D: 69, Baumoiro-2-gil, Seocho-gu, Seoul, 137-900, Korea Tel)+82-2-578-6516, Fax)+82-2-576-2635
Factory: 285-28, Opo-ro, Opo-eup, GwangJu-si, GyeongGi-do, 464-894, Korea Tel)+82-31-766-5974 Fax)+82-31-765-2645
Sales Office: 7, Neunganmal 1 gil, Seocho-gu, Seoul, 137-180, Korea
e-mail: sales@wonatech.com website : www.wonatech.com | www.zivelab.com